

Test certificate

Number **TC6981** revision 0
Project number 600810
Page 1 of 5

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body Number 0122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993 and by application of the OIML International Recommendation R 60 (Edition 2000). The applied error fraction p_j , meant in the paragraph 3.5.4. of the standard is 0.8.

Applicant ARPÈGE MASTER-K
38 Avenue des Frères Montgolfier BP 186
69 686 Chassieu Cedex
France

In respect of A **digital compression load cell**, with strain gauges, tested as a part of a weighing instrument.
Manufacturer : ARPÈGE MASTER-K
Type : DC 285..... and CPFN-A.....

Characteristics

Maximum capacity (E_{max})	15 t up to and including 75 t			
Accuracy class	C			
Maximum number of load cell verification intervals (n_{max})	1000	3000	4000	5000
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	15 000			
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	8000			

In the description number TC6981 revision 0 further characteristics are described.

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI.
B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed.
Parts of the document may only be reproduced after
written permission.

Test certificate

Number **TC6981** revision 0
Project number 600810
Page 2 of 5

Description and documentation The load cell is described in the description number TC6981 revision 0 and documented in the documentation folder TC6981-1, appertaining to this test certificate.

Remarks Summary of the test involved: see Appendix number TC6981 revision 0

Dordrecht, 6 October 2006
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification

Description

Number **TC6981** revision 0
Project number 600810
Page 3 of 5

1 General information about the load cell

All properties of the load cell, whether mentioned or not, may not be in conflict with the standard mentioned in the test certificate.

1.1 Essential parts

Description	Drawing number	Rev.	Remarks
Digi cell 285 Numerique 15 t / 30 t / 50 t / 75 t	103780	1	Mechanical
Digi cell 285 Numerique 15 t / 30 t / 50 t / 75 t	103780A	1	Mechanical
Fulle Bridge circuit type load cell Digi cell 285 15 t / 30 t / 50 t / 75 t	103848	0	Electrical
Implantation Reperes face soudures carte capteur numerique Nouveau Modèle (CCC NM)	104166RES-P0	0	
Implantation Reperes face composants carte capteur numerique Nouveau Modèle (CCC NM)	104166REC-P0	0	
Carte capteur num.nm. proto	104166NOM_PO	0	2 pages

1.2 Essential characteristics

Minimum dead load	: 0 kg
Safe overload	: 150 % of E_{max}
Number of counts for E_{max}	: 100 000,0
Recommended excitation	: 7,5 a 16 V DC
Excitation maximum	: 16 V DC
Transducer material	: Stainless Steel
Atmospheric protection	: Stainless Steel welded IP68

Data transmission:

There are two types of data transmission:

CAN:

- Transmission speed 125 000 bit/s;
- Transmission connection Bus CAN (2wire);
- Cable length 15 m.

RS485:

- Transmission speed 9 600 bit/s;
- Transmission connection RS485 (2 wire);
- 1 Start bit;
- 8 numeric bits;
- 1 Stop bit;
- No Parity bit;
- Cable length 15 m.

Description

Number **TC6981** revision 0
Project number 600810
Page 4 of 5

Calibration procedure:

The calibration procedure is depending of the used indicator.

1,3 Essential shapes

The load cell is built according to drawing:

- Digi cell 285 Numerique 15 t / 30 t / 50 t / 75 t, drawing number 103780;
- Digi cell 285 Numerique 15 t / 30 t / 50 t / 75 t, drawing number 103780A.

The data plate is secured against removal by sealing or will be destroyed when removed. The data plate mentions at least the information and markings as described in the OIML R60 document. In the countries where it is mandatory the load cell should bear this test certificate number: TC6981.

Securing:

The connecting cable of the load cell or the junction box is provided with possibility to seal.

Appendix

Number **TC6981** revision 0
 Project number 600810
 Page 5 of 5

Tests performed for this test certificate:

Test	Institute	type, version, remarks
Temperature test and repeatability (20, 40, -10 and 20 °C)	NMi Certin B.V.	DC285 15 t C4
Temperature effect on minimum dead load output (20, 40, -10 and 20 °C)	NMi Certin B.V.	DC285 15 t C4
Creep (20, 40 and -10 °C)	NMi Certin B.V.	DC285 15 t C4
Minimum dead load output return (20, 40 and -10 °C)	NMi Certin B.V.	DC285 15 t C4
Barometric pressure effects at room temperature	NMi Certin B.V.	DC285 15 t C4
Damp heat, cyclic: marked CH (or not marked)	NMi Certin B.V.	DC285 15 t C4 with washer on the body and DC285 15 t C4 with washer on the top
Additional tests for load cells equipped with electronics:		
Warm-up time	NMi Certin B.V.	DC285 15 t C4
Power voltage variations	NMi Certin B.V.	DC285 15 t C4
Short time power reductions	NMi Certin B.V.	DC285 15 t C4
Bursts (electrical fast transients)	NMi Certin B.V.	DC285 15 t C4
Electrostatic discharge	NMi Certin B.V.	DC285 15 t C4
Electromagnetic susceptibility	NMi Certin B.V.	DC285 15 t C4
Span stability test	NMi Certin B.V.	DC285 15 t C4